

MÁSCARILLA DE PROTECCIÓN INDIVIDUAL

FFP2

DESCRIPCIÓN Y COMPOSICIÓN

- Referencia CTPL-0020
- Fabricada en poliéster
- Forma de cono
- Filtro de partículas micrométricas
- Doble banda elástica de sujeción
- Desechable
- No Estéril
- Libre de látex

CARACTERÍSTICAS

- Tipo de **protección** ≥ 95 (BFE %)
- Máscara **resistente** a la obstrucción en el tiempo (prueba por polvo de Dolomita) que mejora la comodidad de uso.
- **Protección:** Aerosoles sólidos y líquidos (como neblinas de aceite) no tóxicos, de toxicidad media-baja y de alta toxicidad.
- Partículas virales / orgánicas COVID 19
- Fuga hacia el interior total: Satisfactorio
- Contenido de CO₂ en el aire de inhalación: Satisfactorio
- Resistencia a la inhalación: Satisfactorio
- Libre de colofonia

NORMA EU/ESP

UNE-EN 149:2001+A1:2009



ORGANISMO NOTIFICADO
Nº0598

FILTRACIÓN BACTERIANA (BFE) (%)

95

RESPIRABILIDAD (Pa/m²)

28PA/Cm²

EFICACIA DE FILTRACIÓN DE AEROSOL

95%

EFICACIA DE FILTRACIÓN DE PARTICULAS

95%

AJUSTE FACIAL

>95%



INFORME DE ENSAYO / TEST REPORT

2020EC4323

FECHA DE RECEPCIÓN / DATE OF RECEPTION

14/10/2020

FECHA DE ENSAYOS / DATE TESTS

Inicio / Starting: 04/11/2020

Finalización / Ending: 05/11/2020

IDENTIFICACIÓN Y DESCRIPCIÓN DE LAS MUESTRAS / IDENTIFICATION AND DESCRIPTION OF SAMPLES

REFERENCIAS / REFERENCES

MASCARILLA MUNDO SALUD CTPL-0020 CE0598

Descripción /

Description:

Media máscara de partículas sin válvula de exhalación que cubre nariz, boca y mentón, color blanco, de colores

Particle filtering half mask without exhalation valve covering nose, mouth and chin, black colour

ENSAYOS REALIZADOS / TESTS CARRIED OUT

- PENETRACIÓN DEL FILTRO CON CLORURO SÓDICO / FILTER PENETRATION WITH SODIUM CHLORIDE
- RESISTENCIA A LA RESPIRACIÓN / BREATHING RESISTANCE

Rev.1 Esta revisión anula y sustituye a la anterior / *This revision cancels and replaces the previous*
Error por generación incorrecta
Incorrect generation error



RESUMEN DE RESULTADOS / RESULTS SUMMARIZE

ENSAYOS PRELIMINARES PRELIMINARY TESTS

La media máscara referenciada
The half mask referenced

MASCARILLA MUNDO SALUD CTPL-0020 CE0598

En base a los requisitos de las prestaciones de la RfU PPE-R/02.075.02:
Base on the performance requirements of the RfU PPE-R/02.075.02:

https://www.ppe-rfu.eu/app/uploads/sites/10/2020/07/RfU-02.075_02_Regulation.pdf

ENSAYO TEST	REQUISITO REQUIREMENT	RESULTADO RESULT
Penetración del material filtrante con cloruro sódico después de una exposición de 3,5 minutos, tal y como se reciben / <i>Penetration of filtering material with sodium chloride after a 3,5 min exposure, as received</i>	La penetración máxima de la media máscara filtrante, cuando se pruebe con cloruro sódico a 95 l / min, tal y como se reciben, será: <i>The maximum penetration of the filtering half mask, when tested with sodium chloride at 95 l / min, as received, shall be:</i> FFP2, 6% FFP3, 1%	0,78 %
Penetración del material filtrante con cloruro sódico después de una exposición de 120 mg, tal y como se reciben / <i>Penetration of filtering material with sodium chloride after an exposure of 120 mg, as received</i>	La penetración máxima de la media máscara filtrante, cuando se pruebe con cloruro sódico a 95 l/min, después de una exposición de 120 mg, tal y como se reciben, será: <i>The maximum penetration of the filtering half mask with sodium chloride at 95 l/min, after an exposure of 120 mg, as received, shall be:</i> FFP2, 6% FFP3, 1%	---
Resistencia a la respiración, tal y como se reciben / <i>Resistance to breathing, as received</i>	La resistencia máxima para muestras en inhalación a 30 l / min, tal y como se reciben, será: - Para FFP2, resistencia a la respiración a 30 l/min será: 0,7 mbar -Para FFP3, resistencia a la respiración a 30 l/min será: 1,0 mbar. / <i>The maximum resistance for samples in inhalation at 30 l/min, as received, shall be:</i> - For FFP2, the breathing resistance at 30 l/min shall be: 0.7 mbar -For FFP3, the breathing resistance at 30 l/min shall be: 1.0 mbar.	0,47 mbar
	La resistencia máxima para las muestras en inhalación a 95 l / min, tal y como se reciben será: - Para FFP2, resistencia a la respiración a 95 l/min será: 2,4 mbar - Para FFP3, resistencia a la respiración a 95 l/min será: 3,0 mbar / <i>The maximum resistance for samples in inhalation at 95 l/min as received shall be:</i> - For FFP2, the breathing resistance at 95 l/min shall be: 2.4 mbar - For FFP3, the breathing resistance at 95 l/min shall be: 3.0 mbar	1,65 mbar
	La resistencia máxima para las muestras, tal y como se reciben, en exhalación a 160 l/min será de 3,0 mbar. / <i>The maximum resistance for samples, as received, in exhalation at 160 l/min shall be 3.0 mbar</i>	2,69 mbar



RESUMEN DE RESULTADOS / RESULTS SUMMARIZE

ENSAYOS PRELIMINARES PRELIMINARY TESTS

Nota
Remark

- 1- El resto de las pruebas de acuerdo según la norma RfU PPE-R/02.075.02 no han sido evaluadas.
The remaining tests of the according to the standard according to RfU PPE-R/02 075 02 have not been evaluated.
 - 2- Los resultados obtenidos son valores orientativos de un estudio preliminar que no da conformidad al producto.
The obtained results are indicative values of a preliminary study that does not give conformity to the product.
 - 3- Los niveles FFP2 o FFP3 son parámetros según la norma EN 149: 2001 + A1: 2009, no aplicables para RfU PPE-R / 02.075.02 ya que la validez final para el método de certificación COVID 19 es conforme con un PASA / NO PASA.
The levels FFP2 or FFP3 are guidance parameters from the standard EN 149: 2001 + A1. 2009, not applicable for RfU PPE-R/02.075.02 since the final validity for COVID 19 certification method is according to a PASS / NOT PASS.
-

DESCRIPCIÓN DE MUESTRAS / DESCRIPTION OF SAMPLES

FOTOGRAFÍA PHOTOGRAPHY

Referencia Reference

MASCARILLA MUNDO SALUD CTPL-0020 CE0598





RESULTADOS / RESULTS

PENETRACIÓN DEL FILTRO CON CLORURO SÓDICO FILTER PENETRATION WITH SODIUM CHLORIDE

Norma

Standard

EN 149:2001+A1:2009 (punto 8.11) modificada por RfU PPE-R/02.075.02
EN 149:2001+A1:2009 (point 8.11) modified by RfU PPE-R/02.075.02

Aparato

Apparatus

Equipo penetración de cloruro sódico
Sodium chloride penetration equipment

Condiciones de ensayo

Testing conditioning

Fecha de ensayo Test date	Inicial Initial	Final
02/11/2020	21,8 °C / 55,7 %	21,5 °C / 55 %

Pre-acondicionamiento de muestras

Sample Conditioning

- En original
- As received

Observación o desviación de la norma.

Observation or deviation of the standard

Descripción de la muestra

Description of the sample

Media máscara de partículas sin válvula de exhalación que cubre nariz, boca y mentón, color negro.
Particle filtering half mask without exhalation valve covering nose, mouth and chin, black color

Incertidumbre del ensayo

Test uncertainty

La incertidumbre expandida es $\pm 15\%$ del valor del mesurando, para una probabilidad de cobertura del 95%.
The expanded uncertainty is $\pm 15\%$ of the value of the measured for a probability of coverage of 95%.

>>>



RESULTADOS / RESULTS

Referencia

Reference

MASCARILLA MUNDO SALUD CTPL-0020 CE0598

Ensayo de penetración de cloruro sódico (3,5 min) <i>Sodium chloride penetration test (3,5 min)</i>	
Muestra N°. <i>Sample No.</i>	Valor medio de penetración % <i>Average value of penetration %</i>
1	0,33
2	0,78
3	0,29

Exposición a 120 mg de cloruro sódico <i>Exposure to 120mg of sodium chloride</i>	
Muestra N°. <i>Sample No.</i>	Valor Max. De penetración % <i>Max. value of penetration</i>
1	---
2	---
3	---

Requisitos que deben cumplirse según la RfU PPE-R/02.075.02

Requirements to be met according to RfU PPE-R/02.075.02

Máxima penetración de cloruro sódico del material filtrante:

Maximum penetration to the sodium chloride of the filter material:

Clasificación(*) <i>Classification(*)</i>	95l/min %Max
FFP2	6
FFP3	1

(*) Para dar la clasificación se debe de realizar toda la norma EN 149:2001 + A1:2009. Para la Rfu para Covid-19, el requisito es si pasa o no, el ensayo, sin poder clasificar.

() To give the classification, all the EN 149: 2001 + A1: 2009 standard must be carried out. For the Rfu for Covid-19, the requirement is whether or not the test passes, without being able to classify.*

La prueba se llevó a cabo en el laboratorio APPE ubicado en el Polígono Industrial Santiago Payá C/ Filà Benimerines, 25 B - 03801 Alcoy (Alicante).

The test was carried out at APPE Laboratory located at Poligono Industrial Santiago Payá C/ Filà Benimerines, 25 B - 03801 Alcoy (Alicante).

///



RESULTADOS / RESULTS

RESISTENCIA A LA RESPIRACIÓN BREATHING RESISTANCE

Norma

Standard

EN 149:2001+A1:2009 (punto 8.9) modificada por RfU PPE-R/02.075.02
EN 149:2001+A1:2009 (point 8.9) modified by RfU PPE-R/02.075.02

Aparato

Apparatus

Cabezal de prueba Sheffield, equipo de respiración constante y medidor de flujo digital
Sheffield test head, constant breathing equipment and digital flowmeter

Condiciones de ensayo

Testing conditioning

Fecha de ensayo <i>Test date</i>	Inicial <i>Initial</i>	Final
04/11/2020	25 °C / 49,6 %	24,7 °C / 48,5 %

Pre-acondicionamiento de muestras

Sample Conditioning

- En original
- As received

Observación o desviación de la norma.

Observation or deviation of the standard

Descripción de la muestra

Description of the sample

Media máscara de partículas sin válvula de exhalación que cubre nariz, boca y mentón, color negro.
Particle filtering half mask without exhalation valve covering nose, mouth and chin, black color.

Incertidumbre del ensayo

Test uncertainty

La incertidumbre expandida es $\pm 10\%$ del valor del mesurando, para una probabilidad de cobertura del 95%.
The expanded uncertainty is $\pm 10\%$ of the value of the measured for a probability of coverage of 95%.

_____>>>



RESULTADOS / RESULTS

Referencia

Reference

MASCARILLA MUNDO SALUD CTPL-0020 CE0598

Muestra N°. Sample No.	Resistencia a la inhalación Resistance to inhalation (30l/min) mbar	Resistencia a la inhalación Resistance to inhalation (95l/min) mbar
1	0,47	1,65
2	0,35	1,20
3	0,39	1,50

Resistencia a la exhalación (160l/min) mbar Resistance to exhalation (160l/min) mbar					
Muestra N°. Sample No.	Adelante Forward	Hacia arriba Upwards	Abajo Down	Hacia el lado izquierdo Towards the left side	Hacia el lado derecho Towards the right side
1	2,66	2,69	2,69	2,68	2,68
2	2,08	2,09	2,05	2,07	2,07
3	2,44	2,49	2,50	2,48	2,44

Requisitos que deben cumplirse según la RfU PPE-R/02.075.02

Requirements to be met according to RfU PPE-R/02.075.02

Máxima Resistencia permitida (mbar) Maximum resistance permitted (mbar)			
Clasificación(*) Classification(*)	Inhalación 30l/min Inhalation 30l/min	Inhalación 95l/min Inhalation 95l/min	Exhalación 160l/min Exhalation 160l/min
FFP2	0,7	2,4	3,0
FFP3	1,0	3,0	3,0

(*) Los niveles FFP2 o FFP3 son parámetros según la norma EN 149: 2001 + A1: 2009, no aplicables para RfU PPE-R/02.075.02 ya que la validez final para el método de certificación COVID 19 es conforme con un PASA / NO PASA.

(*) The levels FFP2 or FFP3 are guidance parameters from the standard EN 149: 2001 + A1: 2009, not applicable for RfU PPE-R/02.075.02 since the final validity for COVID 19 certification method is according to a PASS / NOT PASS.

La prueba se llevó a cabo en el laboratorio APPE ubicado en el Polígono Industrial Santiago Payá C/ Filà Benimerines, 25 B - 03801 Alcóy (Alicante).

The test was carried out at APPE Laboratory located at Polígono Industrial Santiago Payá C/ Filà Benimerines, 25 B - 03801 Alcoy (Alicante).



Israel Soriano

Responsable Lab. Equipos de Protección Individual Avanzados

Head of Advance Personal Protective Equipment Lab.

CLAUSULAS DE RESPONSABILIDAD

- 1.- AITEX responde únicamente de los resultados sobre los métodos de análisis empleados, consignados en el informe y referidos exclusivamente a los materiales o muestras que se indican en el mismo y que queden en su poder, limitando a éstos la responsabilidad profesional y jurídica del Centro. Salvo mención expresa, las muestras han sido libremente elegidas y enviadas por el solicitante.
- 2.- AITEX no se hace responsable en ningún caso del mal uso de los materiales ensayados ni de la interpretación o uso indebido que pueda hacerse de este documento.
- 3.- La Oferta o Pedido a la que da conformidad el solicitante a través de firma y sello, constituye el Acuerdo Legalmente ejecutable en el que AITEX es responsable de salvaguardar y garantizar, la confidencialidad absoluta, de la gestión de toda la información obtenida o creada durante el desempeño de las actividades contratadas.
- 4.- Ante posibles discrepancias entre informes, se procederá a una comprobación dirimente en la sede central de AITEX. Asimismo, el solicitante se obliga a notificar a AITEX cualquier reclamación que reciba con causa en el informe, eximiendo a este Centro de toda responsabilidad en caso de no hacerlo así, y considerando los plazos de conservación de las muestras.
- 5.- AITEX proporcionará a solicitud del interesado, el procedimiento de tratamiento de quejas.
- 6.- AITEX no se hace responsable de la información proporcionada por los clientes, que se refleja en el Informe, y pueda afectar a la validez de los resultados.
- 7.- AITEX no se hace responsable de un estado inadecuado de la muestra recibida que pudiera comprometer la validez de los resultados, expresando tal circunstancia, en los informes de ensayo.
- 8.- AITEX podrá incluir en sus informes, análisis, resultados, etc., cualquier otra valoración que juzgue necesaria, aún cuando ésta no hubiere sido expresamente solicitada.
- 9.- Cuando se solicite Declaración de Conformidad, de no indicarse lo contrario, se aplicará la regla de decisión según ILAC-G8 & ISO 10576-1 con caso de ambigüedad o indeterminación.
- 10.- Las incertidumbres de ensayos, que se explicitan en el Informe de resultados, se han estimado para una $k=2$ (95% de probabilidad de cobertura). En caso de no informarse, éstas se encuentran a disposición del cliente en AITEX.
- 11.- Los materiales originales, o muestras sobrantes no sometidas a ensayo, se conservarán en AITEX durante los DOCE MESES posteriores a la emisión del informe, por lo que toda comprobación o reclamación que, en su caso, deseará efectuar el solicitante, se deberá ejercer en el plazo indicado.
- 12.- Este informe sólo puede enviarse o entregarse en mano al solicitante o a la persona debidamente autorizada por él.
- 13.- Los resultados de los ensayos y la declaración de cumplimiento con la especificación en este informe se refieren solamente a la muestra de ensayo tal como ha sido analizada/ensayada y no a la muestra/ítem del cual se ha sacado la muestra de ensayo.
- 14.- El cliente debe prestar atención, en todo momento, las fechas de la realización de los ensayos.
- 15.- De acuerdo a la Resolución EA (33) 31, los informes de ensayo deben incluir la identificación única de la muestra pudiendo añadirse además cualquier marca o etiquetado del fabricante. No está permitido reemitir informes de ensayo de denominaciones de muestras (referencias) no ensayadas, sólo se pueden volver a reemitir para la corrección de errores o la inclusión de datos omitidos que ya estaban disponibles en el momento del ensayo. El laboratorio no puede asumir la responsabilidad por la que se declara que el producto con el nuevo nombre comercial / marca comercial es estrictamente idéntico al ensayado originalmente; esta responsabilidad es del cliente.

LIABILITY CLAUSES

- 1.- AITEX is liable only for the results of the methods of analysis used, as expressed in the report and referring exclusively to the materials or samples indicated in the same which are in its possession, the professional and legal liability of the Centre being limited to these. Unless otherwise stated, the samples were freely chosen and sent by the applicant.
- 2.- AITEX shall not be liable in any case of misuse of the test materials nor for undue interpretation or use of this document.
- 3.- The Offer and / or Order to which the applicant gives approval through signature and seal, constitutes the Legally Executable Agreement in which AITEX is responsible for safeguarding and guaranteeing the absolute confidentiality of the management of all the information obtained or created during the performance of the contracted activities.
- 4.- In the eventuality of discrepancies between reports, a check to settle the same will be carried out in the head offices of AITEX. Also, the applicants undertake to notify AITEX of any complaint received by them as a result of the report, exempting this Centre from all liability if such is not done, the periods of conservation of the samples being taken into account.
- 5.- AITEX is not responsible for the information provided by customers, which is reflected in the Report, and may affect the validity of the results.
- 6.- AITEX will provide at the request of the person concerned, the treatment of complaints procedure.
- 7.- AITEX is not responsible for an inadequate state of the sample received that could compromise the validity of the results, expressing such circumstance, in the test reports.
- 8.- AITEX may include in its reports, analyses, results, etc., any other evaluation which it considers necessary, even when it has not been specifically requested.
- 9.- When a Declaration of Conformity is requested, if not indicated otherwise, the decision rule will be applied according to ILAC-G8 & ISO 10576-1, in case of ambiguity, or indeterminacy.
- 10.- The uncertainties of tests, which are made explicit in the Results Report, have been estimated for a $k = 2$ (95% probability of coverage). If not informed, they are available to the client in AITEX.
- 11.- The original materials and rests of samples, not subject to test, will be retained in AITEX during the twelve months following the issuance of the report, so that any check or claim which, in his case, wanted to make the applicant, should be exercised within the period indicated.
- 12.- This report may only be sent or delivered by hand to the applicant or to a person duly authorised by the same.
- 13.- The results of the tests and the statement of compliance with the specification in this report refer only to the test sample as it has been analyzed / tested and not the sample / item which has taken the test sample.
- 14.- The client must attend at all times, to the dates of the realization of the tests.
- 15.- According to Resolution EA (33) 31, the test reports must include the unique identification of the sample, and any brand or label of the manufacturer may be added. It is not allowed to re-issue test reports of untested sample names (references), they can only be re-issued for error correction or inclusion of omitted data that were already available at the time of the test. The laboratory can not assume responsibility for declaring that the product with the new trade name / trademark is strictly identical to the one originally tested; This responsibility belongs to the client.



中国认可
国际互认
检测
TESTING
CNAS L0218



检验检测报告

Test Report



报告编号:	轻委2020-05-0732
委托单位:	东莞康力讯电子科技有限公司 Connexions Technology (Dong Guan) LTD.,
样品名称:	防护口罩 Protective Mask
型号规格:	CTPL-0020
报告日期:	2020年05月27日



广州质量监督检测研究院



广州质量监督检测研究院

Guangzhou Quality Supervision and Testing Institute (GQT)

Page 1 of 3

检验检测报告 Test Report

第 1 页 共 3 页

产品名称 Sample Name 商标 Trademark 型号/规格/等级 Type/Specification/Grade	防护口罩 Protective Mask CTC CTC CTPL-0020 CTPL-0020	生产日期 Production Date 编号或批号 Lot No. 限期使用日期 Expiration Date	2020-05-14 2020-05-14 — — 2年 2years
委托单位 Applicant	东莞康力讯电子科技有限公司 Connexions Technology (Dong Guan) LTD.,	抽样单号 Contract No.	ZJZ200515BYQ-2
生产单位 Manufacturer	东莞康力讯电子科技有限公司 Connexions Technology (Dong Guan) LTD.,	检验类别 Contract Category	委托检验 Applicant Test
来样方式 (抽样者) Sampling Way (Sampler)	委托单位送样 Submit by applicant	样品数量 Sample Quantity	15个 15pieces
样品状况 Condition of Sample	正常 Normal	抽样日期 Sampling Date	2020-05-15
检验环境说明 Test Environment	按标准要求 As the requirements of test reference.	检验日期 Completion Date	2020-05-27
检验依据 Test Reference	中国防护口罩标准《呼吸防护用品 自吸过滤式防颗粒物呼吸器》 Chinese standard Respiratory protective equipment - Non-powered air-purifying particle respirator		
判定依据 Judge Reference	中国防护口罩标准《呼吸防护用品 自吸过滤式防颗粒物呼吸器》 Chinese standard Respiratory protective equipment - Non-powered air-purifying particle respirator		
检验结论 Test Conclusion	所检项目符合中国防护口罩标准要求。 The test items meet the requirements of Chinese standard		
备注 Remarks	此报告以中文为准, 英文仅作参考。 The Chinese version of this test report is the standard one, the English version is only for reference.		



批准:
Approved by

程小烁

审核:
Verified by

李石欣

主检:
Tested by

吴晓兴



(5747/2020.05.28)

地址: 广州市番禺区石楼潮田工业区珠江路1-2号

Address: No.1 Zhujiang Road, Chaotian Industrial Area, Shilou, Panyu District, Guangzhou, Guangdong.

广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 2 of 3

检验检测报告 Test Report

第 2 页 共 3 页

序号 No.	检验项目 Test Items	单位 Unit	标准要求 Reference Requirement	检验结果 Test Results	单项评价 Assess
			— —	— —	
1	过滤效率 Filter efficiency	%	用氯化钠颗粒物检测 样品: ≥ 95.0 Testing by sodium chloride particles Sample: ≥ 95.0	未处理样品 1#: 97.1 2#: 97.4 3#: 96.2 4#: 95.7 5#: 96.2 6#: 97.3 7#: 97.0 8#: 96.8 9#: 97.2 10#: 97.0 温湿度预处理后样品 11#: 96.5 12#: 96.3 13#: 95.3 14#: 95.5 15#: 96.0 Samples as received 1#: 97.1 2#: 97.4 3#: 96.2 4#: 95.7 5#: 96.2 6#: 97.3 7#: 97.0 8#: 96.8 9#: 97.2 10#: 97.0 Samples after temperature and humidity conditioning 11#: 96.5 12#: 96.3 13#: 95.3 14#: 95.5 15#: 96.0	合格 Pass

批准:
Approved by

程小焯

审核:
Verified by

李石欣

主检:
Tested by

吴兴



(5747/2020.05.28)

地址: 广州市番禺区石楼潮田工业区珠江路1-2号

Address: No.1 Zhujiang Road, Chaotian Industrial Area, Shilou, Panyu District, Guangzhou, Guangdong.

广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 3 of 3

检验检测报告 Test Report

第 3 页 共 3 页



批准: 程小炼
Approved by

审核: 李石欣
Verified by

主检: 吴健
Tested by



(5747/2020.05.28)



中国认可
国际互认
检测
TESTING
CNAS L0218



检验检测报告

Test Report



报告编号:	轻委2020-05-0731
委托单位:	东莞康力讯电子科技有限公司 Connexions Technology (Dong Guan) LTD.,
样品名称:	防护口罩 Protective Mask
型号规格:	CTPL-0020
报告日期:	2020年05月27日

广州质量监督检测研究院



广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 1 of 6

检验检测报告 Test Report

第 1 页 共 6 页

产品名称 Sample Name 商标 Trademark 型号/规格/等级 Type/Specification/Grade	防护口罩 Protective Mask CTC CTC CIPL-0020 CIPL-0020	生产日期 Production Date 编号或批号 Lot No. 限期使用日期 Expiration Date	2020-05-14 2020-05-14 — — 2年 2years
委托单位 Applicant	东莞康力讯电子科技有限公司 Connexions Technology (Dong Guan) LTD.,	抽样单号 Contract No.	ZJZ200515BYQ-1
生产单位 Manufacturer	东莞康力讯电子科技有限公司 Connexions Technology (Dong Guan) LTD.,	检验类别 Contract Category	委托检验 Applicant Test
采样方式 (抽送样者) Sampling Way (Sampler)	委托单位送样 Submit by applicant	样品数量 Sample Quantity	20个 20 pieces
样品状况 Condition of Sample	正常 Normal	抽样日期 Sampling Date	2020-05-15
检验环境说明 Test Environment	按标准要求 As the requirements of test reference.	检验日期 Completion Date	2020-05-27
检验依据 Test Reference	中国防护口罩标准《呼吸防护用品 自吸过滤式防颗粒物呼吸器》 Chinese standard Respiratory protective equipment - Non-powered air-purifying particle respirator		
判定依据 Judge Reference	中国防护口罩标准《呼吸防护用品 自吸过滤式防颗粒物呼吸器》 Chinese standard Respiratory protective equipment - Non-powered air-purifying particle respirator		
检验结论 Test Conclusion	所检项目符合中国防护口罩标准要求。 The test items meet the requirements of Chinese standard.		
备注 Remarks	此报告以中文为准,英文仅作参考。 The Chinese version of this test report is the standard one, the English version is only for reference.		

批准:
Approved by

程小烁

审核:
Verified by

李石欣

主检:
Tested by

吴健兴



(5747/2020.05.28)

地址: 广州市番禺区石楼潮田工业区珠江路1-2号

Address: No. 1 Zhujiang Road, Chaotian Industrial Area, Shilou, Panyu District, Guangzhou, Guangdong.

广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 2 of 6

检验检测报告 Test Report

第 2 页 共 6 页

序号 No.	检验项目 Test Items	单位 Unit	标准要求 Reference Requirement	检验结果 Test Results	单项评价 Assess
			— —	— —	
1	一般要求 General requirements	—	材料应满足以下要求: a) 直接与面部接触的材料对皮肤应无害; b) 滤材对人体应无害; c) 所用材料应具有足够的强度, 在正常使用寿命中不应出现破损或变形 The material should meet the following requirements: A) The material directly in contact with the face should be harmless to the skin. B) The filter material should be harmless to the human body. C) The material used should have sufficient strength and should not be damaged or deformed during the period of normal using.	符合标准要求 Meet the requirements	合格 Pass

批准:
Approved by

程小烁

审核:
Verified by

李石欣

主检:
Tested by

吴兴



(5747/2020.05.28)

广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 3 of 6

检验检测报告 Test Report

第 3 页 共 6 页

序号 No.	检验项目 Test Items	单位 Unit	标准要求 Reference Requirement	检验结果 Test Results	单项评价 Assess
			— —	— —	
			结构设计应符合以下要求: a) 应不易产生结构性破损, 部件的设计、组成和安装不应使使用者构成任何危险; b) 头带的设计应可调, 便于佩戴和摘除, 应将面罩牢固地固定在脸上, 且佩戴时不应出现明显的压迫或压痛现象, 可更换式半面罩和全面罩的头带设计应为可更换; c) 应尽可能具有较小的死腔和较大的视野; d) 在佩戴时, 全面罩的镜片不应出现结雾等影响视觉的情况; e) 使用可更换过滤元件、呼气阀、呼气阀以及头罩的呼吸防护用品应采用方便更换的设计, 并且能使使用者随时和方便地检查面罩和面部的佩戴气密性; f) 呼吸导管不应限制头部活动或使用者的行动, 不应影响面罩的密合性, 不应出现限制、阻塞气流的情况; g) 随弃型面罩的结构应能保证与面部的密合, 且应在使用寿命期内不出现变形 The structural or design should meet the following requirements: A) It should not be easy to cause structural damage, and the design, composition and installation of components should not pose any danger to users. B) The design of the head harness should be adjustable, easy to wear and remove. It should be able to fix the mask on the face firmly, and there should be no obvious pressure or tenderness when the mask was worn. The head harness's design of replaceable half face-piece and full face-piece should be replaceable. C) It should have as a smaller dead space and a larger field of vision as possible; D) When worn, the lens of the full face-piece should not appear fogging and other conditions that affect vision. E) Respiratory protective equipment with replaceable filter elements, exhalation valves, inhalation valves and head harness should be designed to be easily replaced, and which can be convenient for the user to check the air tightness between the mask and face at any time. F) Breathing hoses should not restrict head or other movements, should not affect the tightness of the face-piece, and should not restrict or block airflow. G) The structure of the disposable face-piece should be able to ensure the close contact with the face, and there should be no deformation during the period of using.	符合标准要求 Meet the requirements	

批准:
Approved by

程小烁

审核:
Verified by

李石欣

主检:
Tested by

吴晓兴



(5747/2020.05.28)

广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 4 of 6

检验检测报告 Test Report

第 4 页 共 6 页

序号 No.	检验项目 Test Items	单位 Unit	标准要求 Reference Requirement	检验结果 Test Results	单项评价 Assess
			— —	— —	
2	外观 Appearance	—	样品表面不应破损、变形和有明显的其他缺陷。部件材料和结构应能耐受正常使用条件及可能遇到的温度、湿度和机械冲击。头带应可调。可更换式面罩的头带设计应为可更换。全面罩的镜片在佩戴时不应出现结雾等影响视觉的情况。按预处理6.2方法经温度、湿度预处理和机械强度预处理后,部件不应脱落、损坏和变形。检查内容还应包括标识和制造商所提供的各种信息。 The mask should not be damaged, deformed and have obvious other defects. The material and structure of the components should be able to withstand the normal use conditions and the temperature, humidity and mechanical shocks that may be encountered. The head harness should be adjustable and whose design of the replaceable face-piece should be replaceable. The lenses of the full face-piece should not cause fogging and other conditions that affect vision when worn. After temperature and humidity conditioning and mechanical strength conditioning, components should not fall off, be damaged and deformed. The test item also include the identification and various information provided by the manufacturer.	1#: 符合标准要求 2#: 符合标准要求 1#: Meet the requirements 2#: Meet the requirements	合格 Pass
3	呼吸阻力 Breathing resistance	Pa	每个样品的总吸气阻力应不大于350Pa,总呼气阻力应不大于250Pa The total inhalation resistance of each sample should not be more than 350 Pa, and the total exhalation resistance should not be more than 250 Pa.	总吸气阻力: 未处理样品 1#: 78 2#: 77 温度湿度预处理后样品 3#: 66 4#: 69 总呼气阻力: 未处理样品 1#: 82 2#: 82 温度湿度预处理后样品 3#: 65 4#: 75 The total inhalation resistance: Samples as received 1#: 78 2#: 77 Samples after temperature and humidity conditioning 3#: 66 4#: 69 The total exhalation resistance: Samples as received 1#: 82 2#: 82 Samples after temperature and humidity conditioning 3#: 65 4#: 75	合格 Pass

批准:
Approved by

程小烁

审核:
Verified by

李石欣

主检:
Tested by

吴兴



5747/2020.05.28

广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 5 of 6

检验检测报告 Test Report

第 5 页 共 6 页

序号 No.	检验项目 Test Items	单位 Unit	标准要求 Reference Requirement	检验结果 Test Results	单项评价 Assess
			— —	— —	
4	视野 Field of vision	—	半面罩 下方视野: $\geq 60^\circ$ Half face-piece Lower vision: $\geq 60^\circ$	$> 60^\circ$	合格 Pass
5	头带 Head harness	—	随弃式面罩的每条头带、带扣及其他调节部件在承受10N、持续10s的拉力时, 不应出现滑脱或断裂 The head harness, buckle and adjustment components of each disposable face-piece should not be slipped or broken after bearing the pressure of 10N for 10 seconds.	未处理样品 1#: 符合标准要求 温度湿度预处理后样品 2#: 符合标准要求 Sample as received 1#: Meet the requirements Sample after temperature and humidity conditioning 2#: Meet the requirements	合格 Pass

批准:
Approved by

程小烁

审核:
Verified by

李石欣

主检:
Tested by

吴健兴



(5747/2020.05.28)

广州质量监督检测研究院
Guangzhou Quality Supervision and Testing Institute (GQT)

Page 6 of 6

检验检测报告 Test Report

第 6 页 共 6 页



批准: 程小焯
Approved by

审核: 李石欣
Verified by

主检: 吴松兴
Tested by



(5747/2020.05.28)



Certificate F120/966165

Connexions Technology(Dong Guan) LTD.

No.4, Shentan Road, Cishi Town,
DongGuan,
GuangDong,
China.

It is verified that the manufacturer's technical file and the PPE product described on
page 2 have been assessed and found to be in accordance with

Regulation (EU) 2016/425 Module B, EU type-examination

This certificate is valid from 22 July 2020 until 22 July 2025
1. Certified since 22 July 2020

Authorized by

SGS FIMKO OY, Notified Body 0598
Tukimäentie 8, FI-00180, Helsinki, Finland
t +358 9 636 211 f +358 9 632 5474 www.sgs.com

FINAS
Finnish Accreditation Service
S003 (EN ISO/IEC 17065)

Page 1 of 2



This document is issued by the Competent Authority to be the official evidence of
certification services provided by the notified body and its subsidiaries
whenever it refers to the Regulation of the European Union and its
provisions and/or standards. The authority of the document may be verified on
the website www.sgs.com/finland under the section "Notified Bodies".
Any unauthorized alteration, forgery or falsification of the document or any part
of the document is strictly prohibited and offenders may be prosecuted to the fullest
extent of the law.



Certificate FI20/966185, continued

**Connexions Technology(Dong
Guan) LTD.**

Regulation (EU) 2016/425

Module B, EU type-examination

Issue 1

PPE Product

CTC (logo) CTPL-0020 folded half mask, consisting of five
layer (polypropylene/ polypropylene/ polypropylene & hot air
cotton/polyester & polyethylene/ polypropylene) disposable particle filtering
half mask, with nose clip, and nylon/spandex ear band.

It is certified that the manufacturer's technical file and the above mentioned
PPE have been assessed and found to meet the applicable Essential Health
and Safety Requirements in Annex II of Regulation (EU) 2016/425 Personal
Protective Equipment

The following have been applied:

EN 149:2001+A1:2009 (Respiratory protective devices - filtering half masks
to protect against particles) device classification: FFP2 NR.

This certificate is issued on the strict condition that appropriate checks on
manufactured PPE, as detailed in Article 19 (c) of the Regulation are
implemented and maintained while the model is in production

Certification is based on technical file reference:
Face mask/ CTPL-0020, version 2 dated: 2020-07-21

SGS Reference Number UK/CRS 241352

This certificate remains the property of SGS Finke Oy Ltd to whom it must be returned on request.





Fiscal Year 2020

CERTIFICATION OF REGISTRATION

This certifies that:

CONNEXIONS TECHNOLOGY(DONG GUAN)LTD.
No.8, chendian road, Qishi Town, DongGuan, GuangDong China DongGuan,
GUANGDONG, 523000, CHINA

has completed the FDA Establishment Registration and Device Listing

Owner/Operator Number: 10074051

Listing No.	Product Code	Device Name(s)	Premarket Submission Number/Type	Activities
D403004	MSH	Respirator, surgical	Exempt	Manufacturer

Final Registration Date: May 18, 2020
Expiration Date: December 31, 2020



SL52025246081401TX

Date: June 08 2020

Page 3 of 10

Test Result

Respiratory Protective Devices — Filtering Half Masks to Protect against Particles — Requirements, Testing, Marking

EN 149:2001+A1:2009

Clause 7.4 Packaging

(EN 149:2001+A1:2009 Clause 8.2)

Test Requirement	Results	Comment
Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use.	Comply	Pass

Clause 7.5 Material

TEN 148-2001+A1-2009, Clause 5.2 & 5.3.1 & 5.3.2)

Test Requirement	Results	Comment
Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.	Comply	Pass
After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps.	Comply	
When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.	Comply	
Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer.	Comply	

Clause 7.6 Cleaning and Disinfecting

TEN 199-21301-A1-2009, Clause 5.4 & 8.5 & 8.11

Test Requirement	Results	Comment
If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer. With reference to 7.9.2, after cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class.	Not applicable (Not designed to be re-usable)	N.A.

Clause 7.7 Practical Performance

EN 149:2001+A1:2009, Clause 8.4)

Test Requirement	Results	Comment
The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard.	No imperfections	Pass

[illegible]

1. 1999年12月31日	1999年12月31日	1999年12月31日	1999年12月31日
1. 1999年12月31日	1999年12月31日	1999年12月31日	1999年12月31日

© 2007 Blackwell Publishing Ltd *Journal of Internal Medicine* 261: 105–112



Test Report

SL52025245081401TX

Date: June 06, 2020

Page 3 of 10

Clause 7.8 Finish of Parts

(EN 149:2001+A1:2009, Clause 8.2)

Test Requirement	Results	Comment
Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.	No sharp edges or burrs	Pass

Clause 7.9.1 Total Inward Leakage

(EN 149:2001+A1:2009, Clause 8.5)

Test Requirement	Results	Comment
The total inward leakage consists of three components: face seal leakage, exhalation valve leakage (if exhalation valve fitted) and filter penetration. For particle filtering half masks fitted in accordance with the manufacturer's information, at least 48 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than: 25% for FFP1, 11% for FFP2, 5% for FFP3 and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than: 22% for FFP1, 8% for FFP2, 2% for FFP3	Detail refer to Appendix 1	Meet FFP1, Meet FFP2

Appendix 1: Summarization of Test Data**Inward Leakage Test Data**

Subject	Sample No.	Condition	Walk(%)	Head Sideside(%)	Head up/down(%)	Talk(%)	Walk(%)	Mean(%)
Zhou	1	A.R.	6.11	6.06	7.73	6.47	6.33	6.14
Luo	2	A.R.	7.84	6.74	7.17	5.90	8.26	7.18
Lu	3	A.R.	6.44	4.88	6.48	6.18	7.10	6.21
Wang	4	A.R.	5.66	5.60	4.78	5.41	4.82	5.29
Bao	5	A.R.	8.29	8.66	7.51	6.82	5.91	7.44
Ding	6	T.C.	5.59	4.76	5.54	4.97	5.66	5.84
Li	7	T.C.	8.25	7.75	6.81	7.63	8.40	7.77
Chen	8	T.C.	5.45	5.05	5.98	5.58	5.35	5.48
Song	9	T.C.	5.60	6.81	5.64	6.70	6.89	6.35
Ye	10	T.C.	8.33	7.88	8.11	6.80	8.22	7.87

Facial Dimension(mm)

Subject	Face length	Face Width	Face Depth	Mouth Width
Chen	125	150	120	58
Lu	115	132	107	48
Zhou	115	135	106	52
Li	125	130	107	46
Luo	125	136	100	43
Zheng	126	140	112	55
Wang	120	147	103	48
Song	120	140	100	50
Bao	130	134	104	50



This document is signed in writing, representing its issue by the Company under its General Conditions of Service printed below and subject to the provisions of the relevant laws and regulations. The Company's responsibility is limited to the information provided in this document and does not extend to any other information. The Company's responsibility is limited to the information provided in this document and does not extend to any other information. The Company's responsibility is limited to the information provided in this document and does not extend to any other information.

2020年6月6日 14:00:00 14:00:00 14:00:00 14:00:00
14:00:00 14:00:00 14:00:00 14:00:00

Mandate of the NCB Group (SIS 54)



Test Report		SL52025245061401TX	Date: June 08, 2020	Page 4 of 10
Dmg	134	150	110	52
Uu	120	135	117	50
Ve	126	127	105	52

Clause 7.9.2 Penetration of Filter Material
(EN 149:2001+A1:2009, Clause 8.11 & EN 13074-7:2019)

Test Requirement			Results	Comment
The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table			Detail refer to Appendix 2	Meet FFP1, Meet FFP2
Classification	Maximum penetration of test aerosol			
	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min		
	% max.	% max.		
FFP1	20	20		
FFP2	5	5		
FFP3	1	1		

Appendix 2: Summarization of Test Data

Penetration of filter material

Aerosol	Condition	Sample No.	Penetration (%)
Sodium chloride test	As received	1	0.117
		2	0.126
		3	0.119
		4	0.145
	Simulated wearing treatment	5	0.190
		6	0.175
		7	0.412
	Mechanical strength + Temperature conditioned	8	0.317
		9	0.396
Paraffin oil test	As received	10	0.201
		11	0.306
		12	0.287
		13	0.279
	Simulated wearing treatment	14	0.314
		15	0.336
		16	3.674
	Mechanical strength + Temperature conditioned	17	4.124
		18	4.176

Flow conditioning : Single filter: 95.0 L/min



SGS (Swiss Group of Inspection) is a leading global provider of inspection, testing, certification and consulting services. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange.

SGS (Swiss Group of Inspection) is a leading global provider of inspection, testing, certification and consulting services. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange.

SGS (Swiss Group of Inspection) is a leading global provider of inspection, testing, certification and consulting services. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange. SGS is a member of the SGS Group, which is a public limited liability company (PLC) listed on the Swiss Stock Exchange.

MAHSA / PUL SGS Group (8025 SA)



Test Report

SL52025245081401TX

Date: June 08, 2020

Page 6 of 10

Clause 7.13 Head Harness

(EN 149:2001+A1:2009, Clause 8.4 & 9.5)

Test Requirement	Results	Comment
The head harness shall be designed so that the particle filtering half mask can be donned and removed easily.	Comply	Pass
The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device.	Comply	

Clause 7.14 Field of Vision

(EN 149:2001+A1:2009, Clause 8.4)

Test Requirement	Results	Comment
The field of vision is acceptable if determined so in practical performance tests.	Comply	Pass

Clause 7.15 Exhalation Valve(s)

(EN 149:2001+A1:2009, Clause 8.2 & 8.9.1 & 8.3.4 & 8.8)

Test Requirement	Results	Comment
(a) A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations.	Not applicable due to No exhalation valve	N.A.
(b) If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9.	Not applicable due to No exhalation valve	
(c) Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30 s.	Not applicable due to No exhalation valve	
(d) When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10N applied for 10 s.	Not applicable due to No exhalation valve	



SGS (Shanghai) Inspection Co., Ltd. (hereinafter referred to as "SGS") is a member of the SGS Group. SGS is a global leader in inspection, testing, and certification services. SGS is committed to providing high-quality services to its clients and to maintaining the highest standards of integrity and professionalism. SGS is a member of the ISO 9001:2015 standard and is committed to continuous improvement. SGS is a member of the SGS Group and is committed to the same values and standards as the rest of the group. SGS is a member of the SGS Group and is committed to the same values and standards as the rest of the group. SGS is a member of the SGS Group and is committed to the same values and standards as the rest of the group.

SGS (Shanghai) Inspection Co., Ltd. (hereinafter referred to as "SGS") is a member of the SGS Group. SGS is a global leader in inspection, testing, and certification services. SGS is committed to providing high-quality services to its clients and to maintaining the highest standards of integrity and professionalism. SGS is a member of the ISO 9001:2015 standard and is committed to continuous improvement. SGS is a member of the SGS Group and is committed to the same values and standards as the rest of the group. SGS is a member of the SGS Group and is committed to the same values and standards as the rest of the group. SGS is a member of the SGS Group and is committed to the same values and standards as the rest of the group.

Member of the SGS Group (SGS SA)



Test Report

SL52025246081401TX

Date: June 08, 2020

Page 7 of 10

Clause 7.16 Breathing Resistance
(EN 149:2001+A1:2009, Clause 8.9)

Test Requirement				Results	Comment
The penetration of the filter of the particle filtering full mask shall meet the requirements of the following table				Detail refer to Appendix 5	Meq FFP1, Meq FFP2, Meq FFP3
Classification	Maximum permitted resistance (mbar)				
	Inhalation		Exhalation		
	50 l/min	95 l/min	160 l/min		
FFP1	0.6	2.1	3.0		
FFP2	0.7	2.4	3.0		
FFP3	1.0	3.0	3.0		

Appendix 5: Summarization of Test Data

Breathing resistance (mbar)

As received	Flow rate(l/min)	1					2					3				
		A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
		30	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.3
Inhalation	95	0.8	0.7	0.8	0.8	0.8	0.7	0.8	0.8	0.8	0.8	0.7	0.6	0.8	0.8	0.8
	Exhalation	160	1.5	1.6	1.5	1.6	1.7	1.8	1.6	1.6	1.6	1.5	1.6	1.6	1.6	1.6
Simulated wearing treatment	Flow rate(l/min)	4					5					6				
		A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
		30	0.3	0.3	0.2	0.2	0.3	0.3	0.2	0.3	0.3	0.3	0.2	0.3	0.2	0.3
Inhalation	95	0.9	0.8	0.8	0.7	0.8	0.9	0.8	0.8	0.8	0.8	0.9	0.7	0.8	0.8	0.8
	Exhalation	160	1.6	1.6	1.4	1.5	1.6	1.8	1.5	1.5	1.4	1.6	1.8	1.6	1.6	1.5
Temperature conditioned	Flow rate(l/min)	7					8					9				
		A	B	C	D	E	A	B	C	D	E	A	B	C	D	E
		30	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3
Inhalation	95	0.7	0.7	0.6	0.7	0.7	0.8	0.8	0.7	0.6	0.6	0.8	0.6	0.6	0.7	0.7
	Exhalation	160	1.3	1.3	1.4	1.3	1.4	1.3	1.3	1.3	1.4	1.3	1.3	1.4	1.4	1.3

A: facing directly ahead; B: facing vertically upwards; C: facing vertically downwards; D: lying on the left side; E: lying on the right side



Some of the information contained in this document is subject to the Company's internal control system. The Company is not responsible for any loss or damage caused by the use of this document. The Company is not responsible for any loss or damage caused by the use of this document. The Company is not responsible for any loss or damage caused by the use of this document.

Phone: 0086-21-50330000 Fax: 0086-21-50330001
E-mail: info@sgs.com.cn Website: www.sgs.com.cn

Member of the SGS Group (SGS SA)



(EN 149:2001+A1:2009, Clause 8.9 & 8.10)


(EN 149:2001+A1:2009, Clause 8.2)

Test	Uncertainty
Total inward leakage	3.4%
Penetration of filler material	4.6%
Carbon dioxide content of the inhalation air	3.9%
Breathing resistance (30L/min)	5.9%
Breathing resistance (95L/min)	4.9%
Breathing resistance (160L/min)	4.3%



Thellogon (also sold as Thellogon 200)	1.00-2.00M	1.00-2.00M	www.thellogon.com
牛通-100 (牛通100)	1.00-2.00M	1.00-2.00M	www.thellogon.com

Member of the SWS Group (2012-14)





EU DECLARATION OF CONFORMITY

Manufacturer: Connexions Technology(Dong Guan)LTD.,
Address: No.6, Shenbao Road,Qishi Town, DongGuan, GuangDong,China.

declares that the new PPE described hereafter

Product Name	Particle Filtering Half Mask
Model	CTPL-0020
Standard	EN 149:2001+A1:2009
Classification	FFP2 NR

- λ in conformity with the harmonized standard of EN149:2001+A1:2009.
- λ performed the EU type-examination (Module B) of the Regulation (EU) 2016/425 and issued the certificate
- λ performed the conformity to type based on internal production control plus supervised product checks at random intervals (module C2) and issued the certificate .

Notified Body information

Module B	Module C2
SGS Fimko Oy Takomatie 8, FI-00380 Helsinki, Finland Notified Body No. 0598	SGS Fimko Oy Takomatie 8, FI-00380 Helsinki, Finland Notified Body No. 0598

EU REPRESENTATIVE: Power Traders 2020 B.V.

Signed for on behalf of Company

Name: Eric Zhang
Position: General Manager

Date: 2020.11.12

Signature:



DECLARACIÓN DE CONFORMIDAD



MASCARILLA DE PROTECCIÓN INDIVIDUAL FFP2

REFERENCIA DEL FABRICANTE: CTPL-0020

REFERENCIA DEL DISTRIBUIDOR :CTPL-0020 FFP2

Objeto de la siguiente declaración, cumplen con los requisitos establecidos en el REGLAMENTO (UE) 2016/425 DEL PARLAMENTO EUROPEO Y DEL CONSEJO DE LA UNIÓN EUROPEA, de 09 de marzo del 2016, por la que se establece los requisitos sobre el diseño y la fabricación de los equipos de protección individual que vayan a comercializarse, para garantizar la protección de la salud y la seguridad de los usuarios y establecer las normas relativas a la libre circulación de los EPIs en la unión.

ORGANISMO NOTIFICADO:

Nº0598

Cumple con la norma:
UNE EN 149:2001-A1:2009

En Albaurín el Grande a 22 de Julio del 2020